CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 2

CATALOG NUMBER: T-3002

LOT NUMBER: 112014

EXP. DATE: 05-2016

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 2

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 98.0%

COLORMETRIC ANALYSIS

USING HITACHI p-Modular:

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		3.60	3.06 - 4.14	3.60
Specific Gravity		1.017	1.016 - 1.018	1.017
Creatinine	(mg/dL)	3.5	2.98 - 4.03	3.54
Nitrite	(mg/L)	0	0	0
Chromium	(Cr [VI], µg/mL)	0	0	0

Kathy A. Chambers — Quality Control

11/20/2014 — Date

Excel.COA3002.112014

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.