

## **CERTIFICATE OF ANALYSIS**

## **TRUETROL ADULTERATION CONTROL A LEVEL 3**

CATALOG NUMBER:	A-1303 & A-1373	LOT NUMBER: 070523		
		EXP. DATE: 07-31-2024		
PRODUCT NAME:	TEST TRUE TRUETROL CONTROL A LEVEL 3			
APPEARANCE:	CLEAR, COLORLESS SOLUTION			
ANALYTE PURITY	Creatinine	99.0%		
	Sodium Nitrite	99.2%		

COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON OLYMPUS AU 480 ANALYZER. THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIOM RECOMMENDS THAT YOUR FACILITY ESTABLISH IT'S OWN RANGE ON EACH ANALYZER.

<u>ITEM</u>	UNITS	<u>TARGET</u>	SPECIFICATIONS	<u>RESULTS</u>
рН		11.6	10.9 - 12.3	11.5
Specific Gravity		1.001	1.000 - 1.003	1.002
Creatinine	(mg/dL)	23	19.6 - 26.5	23.6
Nitrite	(mg/L)	375	319 - 431	362
Chromium	(Cr [VI], µg/mL)	0	0	0
Oxidant	(mg/L)	395	336 - 454	405
	Terra S. Brown		7/27/2023	
Quality Control		D	ate	

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.