



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 20 mg/dL

CATALOG NUMBER: C-1251 & C-1271 LOT NUMBER: 040725
EXP. DATE: 04-30-2026

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 20 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY CREATININE 99.0%

COLORMETRIC ANALYSIS

USING OLYMPUS AU 480

<u>UNITS</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
mg/dL	20	17 - 23	19.94

Terra S. Brown

Quality Control

4/18/2025

Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.