CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 5 mg/dL

CATALOG NUMBER: C-1252 & C-1272 LOT NUMBER: 020926

EXP. DATE: 2027-08-31

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 5 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine 100.0%

COLORMETRIC ANALYSIS USING OLYMPUS AU 480

UNIT	TARGET	SPECIFICATIONS	RESULTS
mg/dL	5	4.25 - 5.75	5.25

Terra S. Brown 2/26/2026

Quality Control Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.