

CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 100 mg/dL

CATALOG NUMBER: C-1256 & C-1276 LOT NUMBER: 081425

EXP. DATE: 2026-08-31

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 100 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

CREATININE 100.0% **ANALYTE PURITY**

COLORMETRIC ANALYSIS **USING OLYMPUS AU 480**

> **TARGET** UNITS **SPECIFICATIONS RESULTS**

mg/dL 100 85 - 115 106.1

Terra S. Brown 9/2/2025 Date

Quality Control

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.