CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE CALIBRATOR 300 mg/dL

CATALOG NUMBER: C-1258 & C-1278 LOT NUMBER: 111524

EXP. DATE: 11-30-2025

PRODUCT NAME: TEST TRUE CREATININE CALIBRATOR 300 mg/dL

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Creatinine ≥ 98.0%

COLORMETRIC ANALYSIS USING OLYMPUS AU 480	TARGET	SPECIFICATIONS	RESULTS
mg/dL	300	255 - 345	310.31

Terra S. Brown 11/25/2024

Quality Control Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of this product for its particulare use. See product insert for additional indormation.