



## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ NITRITE LINEARITY STANDARD KIT

CATALOG NUMBER: LS5-1010 LOT NUMBER: 062225

EXP. DATE: 2025-12-31

PRODUCT NAME: TEST TRUE NITRITE LINEARITY STANDARD KIT

APPEARANCE: CLEAR, COLORLESS SOLUTIONS

COLORMETRIC ANALYSIS: USING OLYMPUS AU 480

ITEM	TARGET	SPECIFICATIONS	RESULTS
Nitrite	125	106.3 - 143.8	114.31
Nitrite	250	212.5 - 287.5	232.19
Nitrite	500	425 - 575	494.28
Nitrite	1000	850 - 1150	1038.10
Nitrite	1500	1275 - 1725	1630.24

*Terra S. Brown*  
Quality Control

6/24/2025  
Date

---

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.



