



CERTIFICATE OF ANALYSIS

TEST TRUE™ pH LINEARITY STANDARD KIT

CATALOG NUMBER: LS5-1050N & LS5-1050GL LOT NUMBER: 041725

EXP. DATE: 07-31-2025

PRODUCT NAME: TEST TRUE pH LINEARITY STANDARD KIT

APPEARANCE: CLEAR, COLORLESS SOLUTIONS

COLORMETRIC ANALYSIS : USING OLYMPUS AU 480

<u>ITEM</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
pH	2.5	2.6 - 3.0	2.70
pH	3.0	2.7 - 3.3	3.00
pH	4.5	4.1 - 4.9	4.60
pH	5.0	4.6 - 5.4	5.20
pH	7.0	6.6 - 7.4	6.90
pH	9.0	8.5 - 9.5	9.00
pH	10.0	9.5 - 10.5	10.10
pH	11.0	10.5 - 11.5	10.99
pH	11.5	11.2 - 11.8	11.76

Terra S. Brown

Quality Control

4/22/2025

Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.