



CERTIFICATE OF ANALYSIS

TEST TRUE™ CHROMATE LINEARITY STANDARD KIT

CATALOG NUMBER: LS5-1150 & LS6-1150 LOT NUMBER: 102025
EXP. DATE: 2026-04-30

PRODUCT NAME: TEST TRUE CHROMATE LINEARITY STANDARD KIT

APPEARANCE: CLEAR, YELLOW SOLUTIONS

COLORMETRIC ANALYSIS : USING OLYMPUS AU 480

<u>ITEM</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
CHROMATE	10 mg/L	9.0 - 11.0	10.3
CHROMATE	25 mg/L	22.5 - 27.5	25.3
CHROMATE	50 mg/L	45 - 55	50.1
CHROMATE	100 mg/L	90 - 110	97.0
CHROMATE	150 mg/L	135 - 165	142.3
CHROMATE	200 mg/L	180 - 220	186.0

Terra S. Brown
Quality Control

10/21/2025
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.