



CERTIFICATE OF ANALYSIS

TEST TRUE™ CREATININE LINEARITY STANDARD KIT

CATALOG NUMBERS: LS5-1250, LS5-1260, LS5-1270 & LS5-1300 LOT NUMBER: 121325

EXP. DATE: 2026-06-30

PRODUCT NAME: TEST TRUE CREATININE LINEARITY STANDARD KIT

APPEARANCE: CLEAR, COLORLESS SOLUTIONS

COLORMETRIC ANALYSIS : USING OLYMPUS AU 480

<u>ITEM</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
CREATININE	1 mg/dL	.90 - 1.10	1.06
CREATININE	3 mg/dL	2.70 - 3.30	3.07
CREATININE	5 mg/dL	4.50 - 5.50	5.10
CREATININE	10 mg/dL	9.00 - 11.00	10.04
CREATININE	25 mg/dL	22.50 - 27.50	24.99
CREATININE	50 mg/dL	45 - 55	51
CREATININE	100 mg/dL	90 - 110	107
CREATININE	200 mg/dL	180 - 220	199
CREATININE	300 mg/dL	270 - 330	300
CREATININE	400 mg/dL	360 - 440	418
CREATININE	500 mg/dL	450 - 550	513

Terra S. Brown
Quality Control

12/22/2025
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.