



CERTIFICATE OF ANALYSIS

TEST TRUE™ OXIDANT LINEARITY STANDARD KIT

CATALOG NUMBER: LS5-1350 LOT NUMBER: 041925
EXP. DATE: 10-31-2025

PRODUCT NAME: TEST TRUE OXIDANT LINEARITY STANDARD KIT

APPEARANCE: CLEAR, COLORLESS SOLUTIONS

COLORMETRIC ANALYSIS : USING OLYMPUS AU 480

<u>ITEM</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
OXIDANT	37.5 mg/L	31.9 - 43.1	37.4
OXIDANT	75 mg/L	67.5 - 82.5	74.1
OXIDANT	150 mg/L	135 - 165	147.4
OXIDANT	200 mg/L	180 - 220	199.2
OXIDANT	300 mg/L	270 - 330	308.2

Terra S. Brown
Quality Control

4/21/2025
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.