CERTIFICATE OF ANALYSIS

TEST TRUE™ CHROMATE CALIBRATOR 50 mg/L

CATALOG NUMBER: M-1151 & M-1171 LOT NUMBER: 24060520

EXP. DATE: 06-30-2025

PRODUCT NAME: TEST TRUE CHROMATE CALIBRATOR 50 mg/L

ANALYTE PURITY PYRIDINIUM CHLOROCHROMATE 98%

APPEARANCE: CLEAR, YELLOW SOLUTION

COLORMETRIC ANALYSIS USING OLYMPUS AU 480

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	50	42.5 - 57.5	50.5

Terra S. Brown 6/18/2024

Quality Control Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.