



CERTIFICATE OF ANALYSIS

TEST TRUE™ NITRITE CALIBRATOR 1,500 mg/L

CATALOG NUMBER: N-1017 & N-1077 LOT NUMBER: 102623
EXT. EXP. DATE: 10-31-2024

PRODUCT NAME: TEST TRUE NITRITE CALIBRATOR 1,500 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS USING OLYMPUS AU 480

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
Nitrite	mg/L	1,500	1,275 -1,725	1,627.2

Terra S. Brown 10/31/2023
Quality Control Date

Axiom's controls and calibrators are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.