



## CERTIFICATE OF ANALYSIS

---

### TEST TRUE™ NITRITE CALIBRATOR 200 mg/L

CATALOG NUMBER: N-1353 & N-1373 LOT NUMBER: 061325  
EXP. DATE: 2026-06-30

PRODUCT NAME: TEST TRUE NITRITE CALIBRATOR 200 mg/L

ANALYTE PURITY SODIUM NITRITE 99.9%

APPEARANCE: CLEAR, COLORLESS SOLUTION

#### COLORMETRIC ANALYSIS

USING OLYMPUS AU 480

<u>UNITS</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
mg/L	200	170 - 230	204

Terra S. Brown

Quality Control

6/30/2025

Date

---

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.