



CERTIFICATE OF ANALYSIS

TEST TRUE™ OXIDANT (NITRITE) CALIBRATOR 150 mg/L

CATALOG NUMBER: O-1352 & O-1372 LOT NUMBER: 051024
EXT. EXP. DATE: 11-30-2025

PRODUCT NAME: TEST TRUE OXIDANT (NITRITE) CALIBRATOR 150 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

ANALYTE PURITY Sodium Nitrite ≥ 98.0%

COLORMETRIC ANALYSIS

USING OLYMPUS AU 480

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	150	127.5 - 172.5	156.0

Terra S. Brown

Quality Control

5/31/2024

Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.