CERTIFICATE OF ANALYSIS

TEST TRUE™ OXIDANT CALIBRATOR 200 mg/L

CATALOG NUMBER: O-1353 & O-1373 LOT NUMBER: 23060320

EXP. DATE: 06-30-2024

PRODUCT NAME: TEST TRUE OXIDANT CALIBRATOR 200 mg/L

ANALYTE PURITY SODIUM NITRITE 99.9%

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING OLYMPUS AU 480

UNITS	TARGET	SPECIFICATIONS	RESULTS
mg/L	200	170 - 230	209

Terra S. Brown 6/26/2023

Quality Control Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.