CERTIFICATE OF ANALYSIS

TEST TRUE™ pH REAGENT

Catalog Numbers: P-1050N, P-1070/450N, P-1070/100N

LOT NUMBER: 26050110

EXP. DATE: 2028-05-31

Product Name: pH Reagent, R1 (New Formulation)

Appearance: CLEAR, PINKISH WITH GREEN TINT SOLUTION, PARTICULATE FREE

Colormetric Analysis Using Olympus AU480

Total Span (△S2 - △S1) **Specification:** ≥ 0.47 **Result:** 0.5480

Linearity Standard Results: N=5

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH	(units)	3.00	2.6 - 3.5	2.90
pH	(units)	4.50	3.8 - 5.2	4.70
pH	(units)	9.00	7.7 - 10.4	9.20
pH	(units)	10.00	8.5 - 11.5	10.10
pH	(units)	11.00	9.4 - 12.7	11.10

Control Recovery Results: N=10

ITEM	UNITS	TARGET	SPECIFICATIONS	MEAN	CV
pH Contol 1	(units)	3.60	3.1 - 4.1	3.80	0.00%
pH Contol 2	(units)	4.30	3.8 - 4.8	4.60	0.00%
pH Contol 3	(units)	6.00	4.5 - 7.5	5.90	0.00%
pH Contol 4	(units)	10.40	7.8 - 13.0	10.34	0.50%
pH Contol 5	(units)	11.60	8.7 - 14.5	11.76	0.82%
pH Contol 6	(units)	2.70	2.4 - 3.0	2.60	0.00%

CV for all controls must be less than or equal to 5%

Terra S. Brown — Quality Control

05/05/2026 — Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.