



CERTIFICATE OF ANALYSIS

SPECIFIC GRAVITY TEST TRUE CONTROL FOR REFRACTOMETER 1.002

CATALOG NUMBER: R-1125 / R-1225 LOT NUMBER: 120823
EXP. DATE: 06-30-225

PRODUCT NAME: SPECIFIC GRAVITY CONTROL FOR REFRACTOMETER 1.002

APPEARANCE: CLEAR, COLORLESS SOLUTION

REFRACTOMETER ANALYSIS

USING A RUDOLPH MODEL J 57

DIGITAL REFRACTOMETER:

UNITS	TARGET	SPECIFICATIONS	RESULT
Specific Gravity	1.002	1.000 - 1.004	1.0022

Terra S. Brown

Quality Control

12/29/2023

Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.