CERTIFICATE OF ANALYSIS

SPECIFIC GRAVITY TEST TRUE CONTROL FOR REFRACTOMETER 1.011

CATALOG NUMBER: R-1126 / R-1226 LOT NUMBER: 030926

EXP. DATE: 2027-09-30

PRODUCT NAME: SPECIFIC GRAVITY CONTROL FOR REFRACTOMETER 1.011

APPEARANCE: CLEAR, COLORLESS SOLUTION

REFRACTOMETER ANALYSIS

USING RUDOLPH J57 AUTOMATIC DIGITAL REFRACTOMETER

DIGITAL REFRACTOMETER:	TARGET	SPECIFICATIONS	RESULTS
SPECIFIC GRAVITY	1.011	1.009 - 1.013	1.011

Terra S. Brown
Quality Control

3/17/2026
Date

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitabilty of the product for its particular use. See product insert for additional information.