



CERTIFICATE OF ANALYSIS

SPECIFIC GRAVITY TEST TRUE CONTROL FOR REFRACTOMETER 1.023

CATALOG NUMBER: R-1127 / R-1227 LOT NUMBER: 070824
EXP. DATE: 01-31-2026

PRODUCT NAME: SPECIFIC GRAVITY CONTROL FOR REFRACTOMETER 1.023

APPEARANCE: CLEAR, COLORLESS SOLUTION

REFRACTOMETER RESULTS

USING RUDOLPH AUTOMATIC DIGITAL REFRACTOMETER

DIGITAL REFRACTOMETER:	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
Specific Gravity	1.0230	1.0210 - 1.0250	1.0231

Terra S. Brown

Quality Control

7/16/2024

Date

Axiom's controls and calibrators are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.