



CERTIFICATE OF ANALYSIS

TEST TRUE™ SPECIFIC GRAVITY CALIBRATOR 1.020

CATALOG NUMBER: S-1102 & S-1172

LOT NUMBER: 25080720

EXP. DATE: 2026-08-31

PRODUCT NAME: TEST TRUE SPECIFIC GRAVITY 1.020

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS

USING Olympus AU 480

<u>ITEM</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
Specific Gravity	1.020	1.019 - 1.021	1.020

Terra S. Brown

Quality Control

8/19/2025

Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.