CERTIFICATE OF ANALYSIS

TRUETROL ADULTERATION CONTROL LEVEL 7

CATALOG NUMBER: T-3007 & T-3077 LOT NUMBER: 120825

EXP. DATE: 2026-12-31

PRODUCT NAME: TEST TRUE TRUETROL CONTROL LEVEL 7

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON OLYMPUS AU 480 ANALYZER.

THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIOM RECOMMENDS THAT YOUR FACILITY ESTABLISH ITS OWN RANGE ON EACH ANALYZER.

ITEM	UNITS	TARGET	SPECIFICATIONS	RESULTS
pH		4.25	3.85 - 4.65	4.30
Specific Gravity		1.011	1.009 - 1.013	1.010
Creatinine	(mg/dL)	-		
Nitrite	(mg/L)	-		
Chromium	(Cr [VI], µg/mL)	-		

Terra S. Brown — Quality Control

1/19/2026 — Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST traceable standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional indormation.