



## CERTIFICATE OF ANALYSIS

### TRUETROL ADULTERATION CONTROL T LEVEL 8

CATALOG NUMBER: T-3008 & T-3080 LOT NUMBER: 020826  
EXP. DATE: 2027-02-28  
PRODUCT NAME: TEST TRUE TRUETROL CONTROL T LEVEL 8  
APPEARANCE: CLEAR, COLORLESS SOLUTION  
ANALYTE PURITY Creatinine 100.0%  
Sodium Nitrite 99.8%

#### COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON OLYMPUS AU480 ANALYZER.

THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIOM RECOMMENDS THAT YOUR FACILITY ESTABLISH IT'S OWN RANGE ON EACH ANALYZER.

<u>ITEM</u>	<u>UNITS</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
pH		11.6	10.9 - 12.3	11.8
Specific Gravity		1.001	1.000 - 1.003	1.001
Creatinine	(mg/dL)	23.0	19.6 - 26.5	23.7
Nitrite	(mg/L)	250	213 - 288	239
Chromium	(Cr [VI], µg/mL)	0	0	0
Oxidant	(mg/L)	250	213 - 288	256

Terra S. Brown

Quality Control

2/26/2026

Date

Axiom's controls and calibrators are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced against NIST traceable standards.

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.