



## CERTIFICATE OF ANALYSIS

### TRUETROL ADULTERATION CONTROL T LEVEL 13

CATALOG NUMBER: T-3013 & T-3073 LOT NUMBER: 071525  
EXP. DATE: 2026-07-31  
PRODUCT NAME: TEST TRUE TRUETROL CONTROL T LEVEL 13  
APPEARANCE: CLEAR, COLORLESS SOLUTION  
ANALYTE PURITY Creatinine 100.0%  
Sodium Nitrite 99.9%

#### COLORMETRIC ANALYSIS:

VALUES AND RANGES STATED HERE ARE BASED ON OLYMPUS AU 480 ANALYZER.

THE RESULTS OBTAINED ON OTHER ANALYZERS MAY VARY. AXIOM RECOMMENDS THAT YOUR FACILITY ESTABLISH IT'S OWN RANGE ON EACH ANALYZER.

<u>ITEM</u>	<u>UNITS</u>	<u>TARGET</u>	<u>SPECIFICATIONS</u>	<u>RESULTS</u>
pH		6.0	5.4 - 6.6	6.0
Specific Gravity		1.011	1.009 - 1.013	1.010
Creatinine	(mg/dL)	7	6.0 - 8.0	7.2
Nitrite	(mg/L)	250	212.5 - 287.5	238.0
Chromium	(Cr [VI], µg/mL)	0	0	0
Oxidant	(mg/L)	260	221 - 299	269.2

Terra S. Brown

Quality Control

7/22/2025

Date

Axiom's calibrators and controls are prepared to target concentrations by gravimetric methods, with the use of balances that are regularly calibrated and verified with traceability to NIST. Critical components in calibrators and controls are traceable to NIST standards. Automated chemistry analyzers used during QA testing are calibrated with or referenced

Axiom warrants that its products conform to the information contained in this and other Axiom publications. Purchaser must determine the suitability of the product for its particular use. See product insert for additional information.