



CERTIFICATE OF ANALYSIS

TEST TRUE™ ZERO CALIBRATOR 0 mg/L

CATALOG NUMBER: Z-1001 & Z-1071 LOT NUMBER: 081225
EXP. DATE: 2027-02-28

PRODUCT NAME: TEST TRUE ZERO CALIBRATOR 0 mg/L

APPEARANCE: CLEAR, COLORLESS SOLUTION

COLORMETRIC ANALYSIS USING OLYMPUS AU 480:

	TARGET	SPECIFICATIONS	RESULTS
Nitrite	0.00	≤ 1.15	-0.25
Chromate	0.00	≤ 1.15	-0.07
Creatinine	0.00	≤ 1.15	-0.01
Oxidant	0.00	≤ 1.15	0.09

Terra S. Brown

Quality Control

9/2/2025

Date
